

Summary of FY2008 Operational Statistics

Run Identifier	Scheduled User Hours	Storage Ring Availability [hours]	Storage Ring Availability [%]	X-ray Availability [hours]	X-ray Availability [%]	Injector Availability	Faults	Mean Time to Recovery [hours]	Average Fill Duration without a Fault [hours]	Faults Per Day of Delivered Beam	Delivered Integrated Current [A-hr]	Average Current [ma]
Run 07-3	1577.0	1563.1	99.1%	1563.1	99.1%	98.2%	15	0.93	104.2	0.23	154.1	98.6
Run 08-1	1727.0	1672.0	96.8%	1672.0	98.7%	95.8%	13	4.23	128.6	0.19	153.4	91.7
Run 08-2	1284.1	1243.7	96.9%	1243.7	96.9%	95.6%	21	1.93	59.2	0.41	124.3	99.9
Fiscal Year	4588.1	4478.8	97.6%	4478.8	97.6%	98.2%	49	2.23	91.4	0.26	431.7	96.4

Updated

8/18/2008

Data through end of Run 2008-2